



DPD Total Chlorine Reagent Powder Pillows, 25 mL, pk/1000

Продукт №:

1406428

Опасен

BGN Цена:

Свържете се с нас

Налично



For total chlorine determination by the DPD method. USEPA accepted for reporting drinking and wastewater analysis using Hach Method 8167.

Ranges: Low Range: up to - 2.00 mg/L; Mid Range: 0.05 - 4.00 mg/L; High Range: 0 .1- 10.0 mg/L. Sample Size: Low Range - 25 mL. Mid Range: 10 mL. High Range: 5 mL.

Pack of 1000 powder pillows.

Boric Acid free!

Спецификации

EPA compliant:	Да
Бр. тестове:	1000
Измерване, диапазон:	0,02 - 2,00 mg/L Cl ₂
Метод, име:	DPD
Метод, номер:	8167
Параметър:	Total Chlorine, Bromine, Iodine
Платформа:	Прахови вложки
Условия на съхранение:	10 - 25 °C